

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000823  
8:39 30/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:27  
PRESSURIZING 6:17  
STERILIZING 8:05  
136.4 C 222kPa 8:05  
136.1 C 220kPa 8:35  
136.4 C 223kPa 9:05  
136.0 C 220kPa 9:35  
136.4 C 223kPa 10:05  
136.3 C 221kPa 10:35  
136.2 C 222kPa 11:05  
136.3 C 222kPa 11:35  
136.3 C 222kPa 11:35  
VENTING 11:35  
AIR DRYING 12:34  
CYCLE COMPLETE 42:34  
-----

Digital Signature #  
DE165EF74B712FF2  
-----