
Midmark M11 - v1.0.0
Total Cycles: 5170

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2024 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.1 F 0.0
2:00 105.9 F 0.0
4:00 133.9 F 0.1
6:00 162.8 F 0.4
8:00 183.0 F 0.6
10:00 203.4 F 3.2
12:00 222.0 F 6.8
14:00 228.9 F 10.8
16:00 231.5 F 6.8
18:00 241.8 F 11.7
20:00 237.9 F 9.4
22:00 239.1 F 9.3
24:00 249.4 F 14.7
26:00 261.0 F 21.1
28:00 271.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.8
2:00 272.4 F 28.6
3:00 271.7 F 28.0
4:00 271.7 F 28.4

Min 271.3 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 228.5 F 0.0
10:00 226.6 F 0.0
15:00 225.4 F 0.0
20:00 232.1 F 0.0
25:00 228.5 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 28:47
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:07:14
CYCLE COMPLETE

12/06/2024 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____