
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000269
14:35 22/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:01
STERILIZING 3:54
136.4 C 222kPa 3:54
136.4 C 226kPa 4:24
136.0 C 223kPa 4:54
136.4 C 222kPa 5:24
136.4 C 224kPa 5:54
136.4 C 223kPa 6:24
136.4 C 222kPa 6:54
136.4 C 224kPa 7:24
136.4 C 224kPa 7:24
VENTING 7:24
AIR DRYING 8:14
CYCLE COMPLETE 8:32

Digital Signature #
FE4A1DFB4F9B01B0
