
Midmark M11 - v1.0.5
Total Cycles: 4250

Name: RITTER2
Sterilizer ID: V2313228
01/09/2025 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 77.3 F 0.0
4:00 123.4 F 0.0
6:00 114.4 F 0.1
8:00 151.7 F 0.4
10:00 179.8 F 1.1
12:00 200.7 F 4.1
14:00 211.0 F 7.8
16:00 222.8 F 11.7
18:00 226.4 F 6.7
20:00 238.0 F 11.5
22:00 236.1 F 9.1
24:00 239.6 F 10.0
26:00 248.4 F 14.6
28:00 260.4 F 21.3
30:00 271.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.9 F 28.6
2:00 272.4 F 28.9
3:00 272.8 F 29.3
4:00 272.2 F 29.2

Min 271.6 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 216.5 F 0.0
10:00 198.0 F 0.0
15:00 192.8 F 0.0
20:00 192.5 F 0.0
25:00 203.0 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:36
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:08:45
CYCLE COMPLETE

01/09/2025 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____