
Midmark M11 - v1.0.0
Total Cycles: 613

Name: MIDMARK
Sterilizer ID: V2178198
01/20/2021 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.5 F 0.0
2:00 129.7 F 1.0
4:00 156.2 F 3.3
6:00 182.4 F 7.3
8:00 203.5 F 12.3
10:00 220.0 F 5.8
12:00 232.5 F 11.0
14:00 231.2 F 7.3
16:00 243.9 F 13.2
18:00 242.1 F 11.5
20:00 255.6 F 18.4
22:00 267.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.1 F 28.6
2:00 272.4 F 29.0
3:00 272.3 F 29.3
4:00 271.5 F 28.4

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.0 F 0.0
5:00 228.1 F 0.0
10:00 240.2 F 0.0
15:00 226.4 F 0.0
20:00 221.9 F 0.0
25:00 202.3 F 0.0
Dry Cycle Aborted

FILLING: 0:57
HEATING: 23:18
STERILIZING: 4:00
VENTING: 3:02
DRYING: 25:32
TOTAL CYCLE: 00:56:49

01/20/2021 02:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____