
Midmark M11 - v1.0.0
Total Cycles: 2699

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2022 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 154.7 F 0.9
4:00 177.6 F 3.2
6:00 193.8 F 7.0
8:00 214.0 F 12.6
10:00 222.4 F 6.9
12:00 235.6 F 12.2
14:00 236.1 F 9.2
16:00 238.3 F 9.2
18:00 248.8 F 14.4
20:00 260.8 F 21.3
22:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.3 F 28.7
2:00 271.4 F 28.3
3:00 272.5 F 28.8
4:00 271.4 F 28.1

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 208.4 F 0.0
10:00 227.7 F 0.0
15:00 230.0 F 0.0
20:00 225.6 F 0.0
Dry Cycle Aborted

FILLING: 0:51
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:11
DRYING: 21:16
TOTAL CYCLE: 00:51:45

12/20/2022 01:58 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____