
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002536
8:56 24/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:05
PRESSURIZING 1:10
STERILIZING 2:13
136.0 C 220kPa 2:13
136.4 C 224kPa 2:43
136.4 C 225kPa 3:13
136.3 C 223kPa 3:43
136.4 C 222kPa 4:13
136.2 C 223kPa 4:43
136.4 C 225kPa 5:13
136.2 C 223kPa 5:43
136.2 C 223kPa 5:43
VENTING 5:43
AIR DRYING 6:31
CYCLE COMPLETE 36:31

Digital Signature #
82CD8A10E4CDABB7
