
Midmark M11 - v1.0.5
Total Cycles: 2664

Name: RITTER2
Sterilizer ID: V2313228
07/18/2023 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.0 F 0.0
2:00 199.1 F 4.0
4:00 205.3 F 9.6
6:00 222.8 F 6.9
8:00 233.6 F 11.3
10:00 234.9 F 9.4
12:00 237.5 F 9.3
14:00 249.6 F 15.6
16:00 262.0 F 22.7
18:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.4 F 29.4
2:00 271.9 F 29.6
3:00 272.2 F 29.3
4:00 272.2 F 29.7

Min 271.5 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 213.0 F 0.0
10:00 220.1 F 0.0
15:00 218.2 F 0.0
20:00 218.2 F 0.0
25:00 218.8 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 2:32
HEATING: 18:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

07/18/2023 02:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____