

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000337  
12:16 08/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:04  
STERILIZING 4:18  
136.0 C 229kPa 4:18  
136.2 C 231kPa 4:48  
136.4 C 228kPa 5:18  
136.3 C 232kPa 5:48  
136.4 C 231kPa 6:18  
136.4 C 233kPa 6:48  
136.4 C 232kPa 7:18  
136.4 C 232kPa 7:48  
136.4 C 232kPa 7:48  
VENTING 7:48  
CYCLE ABORTED 8:11  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 8:11  
-----

Digital Signature #  
2FCDE75DF553802F  
-----