
Midmark M11 - v1.0.0
Total Cycles: 4824

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2024 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 73.2 F 0.0
4:00 78.7 F 0.0
6:00 101.8 F 0.1
8:00 131.2 F 0.5
10:00 167.1 F 1.5
12:00 192.8 F 3.5
14:00 208.4 F 7.1
16:00 217.1 F 10.6
18:00 227.8 F 7.5
20:00 233.4 F 8.8
22:00 243.3 F 13.4
24:00 240.3 F 10.7
26:00 235.1 F 8.3
28:00 249.6 F 14.8
30:00 260.5 F 21.0
32:00 270.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.7 F 28.9
2:00 272.0 F 28.3
3:00 271.4 F 28.3
4:00 272.8 F 28.9

Min 271.2 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 218.9 F 0.0
10:00 197.2 F 0.0
15:00 247.3 F 0.0
20:00 233.9 F 0.0
25:00 231.9 F 0.0
30:00 227.9 F 0.0

Drying Complete

FILLING: 0:40
HEATING: 32:52
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:10:58
CYCLE COMPLETE

08/16/2024 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____