
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008284
11:55 31/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:28
PRESSURIZING 1:32
STERILIZING 2:33
136.4 C 220kPa 2:33
136.4 C 224kPa 3:03
136.4 C 221kPa 3:33
136.4 C 223kPa 4:03
136.2 C 217kPa 4:33
136.4 C 222kPa 5:03
136.4 C 221kPa 5:33
136.5 C 224kPa 6:03
136.5 C 224kPa 6:03
VENTING 6:03
AIR DRYING 6:53
CYCLE COMPLETE 17:13

Digital Signature #
57EC1B9706C63F45
