
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000932
12:38 16/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:35
STERILIZING 3:48
136.0 C 224kPa 3:48
136.4 C 232kPa 4:18
136.4 C 231kPa 4:48
136.3 C 231kPa 5:18
136.4 C 232kPa 5:48
136.4 C 229kPa 6:18
136.2 C 230kPa 6:48
136.4 C 232kPa 7:18
136.4 C 232kPa 7:18
VENTING 7:18
AIR DRYING 8:11
CYCLE COMPLETE 23:11

Digital Signature #
9440435B3B37486A
