

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2282

Name: RITTER 1  
Sterilizer ID: V2178198  
09/02/2022 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.0 F 0.0  
2:00 128.9 F 0.5  
4:00 162.3 F 2.8  
6:00 188.7 F 7.5  
8:00 206.9 F 13.3  
10:00 220.4 F 6.9  
12:00 233.9 F 11.8  
14:00 233.4 F 8.1  
16:00 243.0 F 11.7  
18:00 244.5 F 12.3  
20:00 257.5 F 19.1  
22:00 268.8 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.7  
2:00 271.5 F 28.0  
3:00 272.8 F 28.8  
4:00 271.8 F 28.0

Min 271.5 F 27.9  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 214.9 F 0.0  
10:00 216.9 F 0.0  
15:00 209.9 F 0.0  
20:00 204.4 F 0.0  
25:00 202.0 F 0.0  
30:00 197.0 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:08  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:00:58  
CYCLE COMPLETE

09/02/2022 12:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_