
Midmark M11 - v1.0.5
Total Cycles: 2394

Name: RITTER2
Sterilizer ID: V2313228
05/19/2023 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.3 F 0.0
2:00 165.1 F 2.5
4:00 171.4 F 5.8
6:00 193.9 F 10.5
8:00 214.2 F 4.6
10:00 226.9 F 9.8
12:00 232.3 F 8.4
14:00 242.0 F 12.9
16:00 241.6 F 11.6
18:00 254.2 F 18.2
20:00 265.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.2
1:00 272.6 F 30.0
2:00 272.1 F 29.3
3:00 272.2 F 29.9
4:00 272.5 F 29.6

Min 271.6 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 205.4 F 0.0
10:00 234.2 F 0.0
15:00 230.8 F 0.0
20:00 220.7 F 0.0
25:00 217.4 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

05/19/2023 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____