

-----  
Midmark M11 - v1.0.0  
Total Cycles: 666

Name: MIDMARK  
Sterilizer ID: V2178198  
02/11/2021 02:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.0 F 0.0  
2:00 107.6 F 0.0  
4:00 115.2 F 0.1  
6:00 143.3 F 0.8  
8:00 174.7 F 3.8  
10:00 194.0 F 7.2  
12:00 214.5 F 11.0  
14:00 228.3 F 5.3  
16:00 234.8 F 9.7  
18:00 236.7 F 9.4  
20:00 243.7 F 11.9  
22:00 240.2 F 10.0  
24:00 252.8 F 16.3  
26:00 263.9 F 22.7  
28:00 271.6 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.2 F 28.2  
2:00 272.6 F 28.5  
3:00 272.7 F 28.8  
4:00 272.0 F 28.6  
  
Min 271.7 F 28.2  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 249.0 F 0.0  
10:00 230.5 F 0.0  
15:00 230.0 F 0.0  
20:00 226.1 F 0.0  
25:00 228.8 F 0.0  
30:00 232.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 28:47  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:07:02  
CYCLE COMPLETE

02/11/2021 03:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_