

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3348

Name: RITTER2  
Sterilizer ID: V2313228  
02/02/2024 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 88.1 F 0.0  
4:00 112.6 F 0.0  
6:00 110.1 F 1.0  
8:00 144.9 F 3.7  
10:00 180.4 F 7.4  
12:00 203.2 F 11.2  
14:00 218.3 F 4.4  
16:00 225.4 F 7.4  
18:00 236.3 F 11.4  
20:00 230.1 F 7.2  
22:00 242.0 F 12.5  
24:00 239.9 F 10.8  
26:00 252.9 F 17.2  
28:00 264.0 F 23.7  
30:00 271.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.6 F 29.5  
2:00 271.3 F 28.9  
3:00 272.8 F 29.6  
4:00 271.5 F 28.8

Min 271.3 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
5:00 233.9 F 0.0  
10:00 229.9 F 0.0  
15:00 222.9 F 0.0  
20:00 229.9 F 0.0  
25:00 226.9 F 0.0  
30:00 218.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:06  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:08:09  
CYCLE COMPLETE

02/02/2024 10:03 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_