
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001627
16:38 27/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:21
STERILIZING 3:24
136.0 C 222kPa 3:24
136.2 C 224kPa 3:54
136.3 C 225kPa 4:24
136.4 C 226kPa 4:54
136.1 C 223kPa 5:24
136.3 C 224kPa 5:54
136.4 C 225kPa 6:24
136.4 C 225kPa 6:54
136.4 C 225kPa 6:54
VENTING 6:54
AIR DRYING 7:45
CYCLE COMPLETE 7:55

Digital Signature #
47B22268741B57E8
