
Midmark M11 - v1.0.0
Total Cycles: 1531

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2022 10:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.3 F 0.0
2:00 120.8 F 0.0
4:00 134.0 F 2.3
6:00 159.3 F 5.6
8:00 184.4 F 9.7
10:00 205.2 F 6.3
12:00 216.6 F 6.7
14:00 228.4 F 11.0
16:00 229.8 F 7.0
18:00 242.7 F 12.8
20:00 240.7 F 11.0
22:00 253.0 F 17.2
24:00 264.1 F 23.7
26:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.3
2:00 271.5 F 28.6
3:00 272.5 F 29.5
4:00 272.0 F 28.9

Min 271.4 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 211.4 F 0.0
10:00 225.3 F 0.0
15:00 230.1 F 0.0
20:00 223.2 F 0.0
25:00 218.8 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:05
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:04:16
CYCLE COMPLETE

01/17/2022 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____