
Midmark M11 - v1.0.0
Total Cycles: 1437

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2021 05:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 197.2 F 3.6
4:00 210.5 F 9.7
6:00 224.0 F 6.0
8:00 238.1 F 11.6
10:00 236.6 F 9.2
12:00 236.9 F 8.5
14:00 250.6 F 15.4
16:00 263.6 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.7 F 28.1
2:00 272.2 F 28.5
3:00 272.6 F 29.0
4:00 271.5 F 28.2

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 235.3 F 0.0
10:00 226.7 F 0.0
15:00 226.2 F 0.0
20:00 223.7 F 0.0
25:00 221.6 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 17:49
STERILIZING: 4:00
VENTING: 2:47
DRYING: 30:00
TOTAL CYCLE: 00:57:01
CYCLE COMPLETE

12/09/2021 07:44 AM
ACCEPTED: X
REJECTED:
Operator: U02