
Midmark M11 - v1.0.5
Total Cycles: 495

Name: RITTER2
Sterilizer ID: V2313228
10/28/2021 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.2 F 0.0
2:00 135.2 F 1.3
4:00 141.2 F 3.7
6:00 170.5 F 7.7
8:00 191.3 F 12.7
10:00 217.1 F 7.7
12:00 235.9 F 13.0
14:00 237.3 F 9.8
16:00 234.1 F 9.0
18:00 246.7 F 13.8
20:00 258.3 F 20.2
22:00 269.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.9
1:00 272.0 F 30.2
2:00 272.1 F 30.2
3:00 272.1 F 30.1
4:00 272.0 F 30.0

Min 271.7 F 29.8
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 224.1 F 0.0
10:00 223.5 F 0.0
15:00 217.2 F 0.0
20:00 210.9 F 0.0
25:00 210.7 F 0.0
30:00 207.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 23:52
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:02:02
CYCLE COMPLETE

10/28/2021 09:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____