
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001052
14:37 29/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:01
STERILIZING 3:10
136.4 C 225kPa 3:10
136.5 C 226kPa 3:40
136.4 C 224kPa 4:10
136.3 C 222kPa 4:40
136.4 C 225kPa 5:10
136.4 C 225kPa 5:40
136.4 C 223kPa 6:10
136.4 C 225kPa 6:40
136.4 C 225kPa 6:40
VENTING 6:40
AIR DRYING 7:31
CYCLE COMPLETE 22:31

Digital Signature #
1D09D86C72669B4B
