

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3755

Name: RITTER2  
Sterilizer ID: V2313228  
07/05/2024 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.8 F 0.0  
2:00 115.2 F 0.8  
4:00 121.3 F 2.6  
6:00 144.3 F 5.4  
8:00 174.2 F 8.7  
10:00 192.8 F 12.7  
12:00 210.9 F 5.2  
14:00 221.0 F 8.8  
16:00 230.8 F 12.7  
18:00 232.2 F 9.0  
20:00 240.5 F 11.5  
22:00 242.7 F 12.5  
24:00 253.9 F 18.2  
26:00 264.2 F 24.3  
28:00 272.2 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.3  
1:00 272.4 F 30.1  
2:00 272.1 F 29.6  
3:00 271.0 F 29.2  
4:00 271.7 F 30.1

Min 270.9 F 29.1  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 197.8 F 0.0  
10:00 195.9 F 0.0  
15:00 191.9 F 0.0  
20:00 190.3 F 0.0  
25:00 191.2 F 0.0  
30:00 190.5 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:14  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:06:56  
CYCLE COMPLETE

07/05/2024 11:13 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_