

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1731

Name: RITTER2  
Sterilizer ID: V2313228  
11/16/2022 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
2:00 175.0 F 0.3  
4:00 191.9 F 1.1  
6:00 211.7 F 6.0  
8:00 223.7 F 11.4  
10:00 230.4 F 8.3  
12:00 239.2 F 10.9  
14:00 245.1 F 13.5  
16:00 245.5 F 13.2  
18:00 258.6 F 20.5  
20:00 270.3 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.5  
2:00 271.6 F 29.0  
3:00 272.5 F 29.5  
4:00 271.6 F 28.8

Min 271.5 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.7 F 0.0  
5:00 217.6 F 0.0  
10:00 236.7 F 0.0  
15:00 245.4 F 0.0  
20:00 239.1 F 0.0  
25:00 223.7 F 0.0  
30:00 211.4 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 20:50  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:03  
CYCLE COMPLETE

11/16/2022 11:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_