
Midmark M11 - v1.0.5
Total Cycles: 3853

Name: RITTER2
Sterilizer ID: V2313228
08/13/2024 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 150.3 F 1.7
4:00 162.8 F 4.8
6:00 191.6 F 9.1
8:00 208.6 F 10.6
10:00 223.4 F 8.7
12:00 232.8 F 11.7
14:00 238.1 F 11.1
16:00 237.0 F 9.4
18:00 250.4 F 15.8
20:00 260.9 F 22.3
22:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.6
1:00 271.3 F 29.5
2:00 272.3 F 29.7
3:00 272.9 F 30.5
4:00 272.5 F 30.4

Min 271.0 F 29.5
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 237.9 F 0.0
10:00 208.3 F 0.0
15:00 179.8 F 0.0
20:00 174.8 F 0.0
25:00 171.6 F 0.0
30:00 169.2 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

08/13/2024 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____