
Midmark M11 - v1.0.0
Total Cycles: 3133

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2023 07:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.2 F 0.0
2:00 85.2 F 0.0
4:00 99.3 F 0.0
6:00 117.2 F 0.0
8:00 156.4 F 0.0
10:00 193.9 F 0.8
12:00 203.4 F 4.4
14:00 219.8 F 7.4
16:00 232.1 F 10.7
18:00 231.8 F 7.7
20:00 240.0 F 10.5
22:00 234.9 F 8.0
24:00 247.2 F 13.5
26:00 244.6 F 12.1
28:00 256.8 F 18.5
30:00 267.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.8
1:00 272.1 F 28.5
2:00 271.9 F 28.0
3:00 271.8 F 28.3
4:00 272.3 F 28.3

Min 271.3 F 27.6
Max 272.6 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 197.8 F 0.0
10:00 217.3 F 0.0
15:00 224.7 F 0.0
20:00 227.2 F 0.0
25:00 235.3 F 0.0
30:00 233.7 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:21
CYCLE COMPLETE

04/20/2023 08:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____