
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000085
11:44 21/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:27
STERILIZING 4:33
136.0 C 221kPa 4:33
136.1 C 223kPa 5:03
136.2 C 223kPa 5:33
136.4 C 225kPa 6:03
136.4 C 223kPa 6:33
136.4 C 224kPa 7:03
136.0 C 222kPa 7:33
136.4 C 223kPa 8:03
136.4 C 223kPa 8:03
VENTING 8:03
AIR DRYING 8:59
CYCLE COMPLETE 12:20

Digital Signature #
B3FA91B1B7EF7EC5
