
Midmark M11 - v1.0.5
Total Cycles: 3905

Name: RITTER2
Sterilizer ID: V2313228
09/04/2024 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 77.6 F 0.0
4:00 84.0 F 0.0
6:00 103.6 F 0.1
8:00 154.9 F 0.4
10:00 182.8 F 2.3
12:00 198.8 F 5.6
14:00 207.3 F 9.2
16:00 219.6 F 12.8
18:00 224.8 F 7.2
20:00 235.9 F 11.5
22:00 233.7 F 8.4
24:00 242.7 F 11.8
26:00 244.9 F 12.8
28:00 256.5 F 19.0
30:00 267.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.2 F 28.9
2:00 272.5 F 29.3
3:00 271.6 F 29.0
4:00 271.8 F 28.7

Min 271.4 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 222.1 F 0.0
10:00 187.5 F 0.0
15:00 188.3 F 0.0
20:00 188.1 F 0.0
25:00 186.7 F 0.0
30:00 184.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

09/04/2024 10:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____