
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001422
11:09 14/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:31
STERILIZING 4:45
136.0 C 220kPa 4:45
136.4 C 223kPa 5:15
136.4 C 224kPa 5:45
136.3 C 222kPa 6:15
136.2 C 222kPa 6:45
136.3 C 221kPa 7:15
136.1 C 221kPa 7:45
136.4 C 224kPa 8:15
136.4 C 224kPa 8:15
VENTING 8:15
AIR DRYING 9:04
CYCLE COMPLETE 24:05

Digital Signature #
0C299BDD0E21BE50
