
Midmark M11 - v1.0.5
Total Cycles: 3908

Name: RITTER2
Sterilizer ID: V2313228
09/05/2024 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.5 F 0.0
2:00 102.1 F 0.0
4:00 129.4 F 0.3
6:00 149.2 F 1.6
8:00 173.9 F 4.7
10:00 181.8 F 8.2
12:00 197.9 F 11.9
14:00 215.5 F 5.1
16:00 224.3 F 9.2
18:00 234.9 F 12.8
20:00 237.9 F 10.6
22:00 237.2 F 8.6
24:00 248.1 F 14.7
26:00 258.9 F 20.6
28:00 268.8 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.6 F 29.8
2:00 272.4 F 30.1
3:00 272.3 F 29.5
4:00 272.8 F 30.1

Min 271.8 F 29.4
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 212.8 F 0.0
10:00 209.4 F 0.0
15:00 211.7 F 0.0
20:00 213.5 F 0.0
25:00 220.6 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:59
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

09/05/2024 12:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____