
Midmark M11 - v1.0.5
Total Cycles: 3769

Name: RITTER2
Sterilizer ID: V2313228
07/11/2024 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.7 F 0.0
2:00 152.4 F 1.9
4:00 163.8 F 5.1
6:00 187.9 F 9.3
8:00 209.5 F 9.1
10:00 222.4 F 8.4
12:00 236.3 F 12.9
14:00 236.4 F 10.8
16:00 236.2 F 9.2
18:00 250.1 F 15.8
20:00 261.4 F 22.3
22:00 271.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.8 F 30.2
2:00 271.9 F 30.2
3:00 271.4 F 29.7
4:00 272.1 F 29.8

Min 270.9 F 29.1
Max 273.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 244.2 F 0.0
10:00 220.6 F 0.0
15:00 224.2 F 0.0
20:00 224.1 F 0.0
25:00 233.0 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

07/11/2024 12:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____