
Midmark M11 - v1.0.0
Total Cycles: 4408

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2024 07:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 84.7 F 0.1
4:00 82.3 F 0.2
6:00 111.4 F 1.4
8:00 145.4 F 3.8
10:00 163.6 F 7.1
12:00 185.6 F 10.7
14:00 205.4 F 5.4
16:00 213.6 F 4.8
18:00 223.3 F 8.0
20:00 233.2 F 11.7
22:00 231.3 F 7.3
24:00 243.2 F 12.4
26:00 239.9 F 10.3
28:00 252.2 F 16.3
30:00 262.9 F 22.4
32:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.4 F 28.9
2:00 271.5 F 28.0
3:00 272.7 F 28.9
4:00 271.5 F 28.1

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 226.8 F 0.0
10:00 225.2 F 0.0
15:00 228.0 F 0.0
20:00 229.7 F 0.0
25:00 229.3 F 0.0
30:00 228.0 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:12
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:28
CYCLE COMPLETE

03/27/2024 08:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____