
Midmark M11 - v1.0.5
Total Cycles: 220

Name: RITTER2
Sterilizer ID: V2313228
07/19/2021 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 75.5 F 0.0
4:00 96.3 F 0.1
6:00 101.8 F 1.1
8:00 125.7 F 3.7
10:00 155.2 F 6.9
12:00 179.0 F 10.4
14:00 195.3 F 8.6
16:00 208.1 F 6.5
18:00 221.4 F 10.2
20:00 230.6 F 9.7
22:00 234.2 F 11.2
24:00 233.5 F 8.7
26:00 245.8 F 14.5
28:00 256.8 F 20.4
30:00 266.6 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.9
1:00 271.8 F 30.5
2:00 272.3 F 30.3
3:00 271.3 F 29.9
4:00 272.3 F 30.8

Min 271.2 F 29.8
Max 272.9 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 210.5 F 0.0
10:00 195.4 F 0.0
15:00 203.2 F 0.0
20:00 195.5 F 0.0
25:00 190.1 F 0.0
Dry Cycle Aborted

FILLING: 0:50
HEATING: 31:34
STERILIZING: 4:00
VENTING: 3:11
DRYING: 28:24
TOTAL CYCLE: 01:07:60

07/19/2021 09:13 AM
ACCEPTED: X
REJECTED:
Operator: U01