

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008300  
9:07 10/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:30  
STERILIZING 3:35  
136.4 C 220kPa 3:35  
136.3 C 223kPa 4:05  
136.4 C 221kPa 4:35  
136.4 C 220kPa 5:05  
136.3 C 222kPa 5:35  
136.3 C 222kPa 6:05  
136.4 C 223kPa 6:35  
136.4 C 221kPa 7:05  
136.4 C 221kPa 7:05  
VENTING 7:06  
AIR DRYING 7:54  
CYCLE COMPLETE 37:55

-----  
Digital Signature #  
AE7D7414343C9CAF  
-----