
Midmark M11 - v1.0.5
Total Cycles: 661

Name: RITTER2
Sterilizer ID: V2313228
12/29/2021 08:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.5 F 0.0
2:00 75.7 F 0.0
4:00 84.3 F 0.0
6:00 103.5 F 1.7
8:00 138.7 F 4.5
10:00 167.8 F 7.9
12:00 183.8 F 11.6
14:00 207.1 F 4.6
16:00 215.4 F 8.6
18:00 227.2 F 12.6
20:00 229.4 F 8.7
22:00 241.4 F 12.9
24:00 240.8 F 11.7
26:00 252.6 F 17.6
28:00 263.1 F 23.8
30:00 272.9 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.9
1:00 272.5 F 30.0
2:00 272.7 F 30.4
3:00 272.2 F 30.3
4:00 272.2 F 29.8

Min 271.8 F 29.8
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 209.8 F 0.0
10:00 216.8 F 0.0
15:00 207.9 F 0.0
20:00 206.5 F 0.0
25:00 210.8 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:10
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

12/29/2021 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____