
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006402
10:11 04/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:21
STERILIZING 4:25
136.5 C 223kPa 4:25
136.4 C 225kPa 4:55
136.4 C 222kPa 5:25
136.4 C 220kPa 5:55
136.1 C 222kPa 6:25
136.4 C 224kPa 6:55
136.4 C 223kPa 7:25
136.4 C 218kPa 7:55
136.4 C 218kPa 7:55
VENTING 7:55
AIR DRYING 8:47
CYCLE COMPLETE 9:40

Digital Signature #
7108C785D848A1E8
