
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008041
8:45 08/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:09
STERILIZING 4:11
136.0 C 220kPa 4:11
136.4 C 221kPa 4:41
136.3 C 222kPa 5:11
136.4 C 221kPa 5:41
136.4 C 219kPa 6:11
136.4 C 219kPa 6:41
136.3 C 219kPa 7:11
136.4 C 219kPa 7:41
136.4 C 219kPa 7:41
VENTING 7:41
AIR DRYING 8:40
CYCLE COMPLETE 25:42

Digital Signature #
BED6E56236785E9E
