
Midmark M11 - v1.0.5
Total Cycles: 2303

Name: RITTER2
Sterilizer ID: V2313228
04/27/2023 07:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 80.6 F 0.0
4:00 111.9 F 0.0
6:00 126.0 F 0.0
8:00 151.5 F 2.2
10:00 179.6 F 5.1
12:00 190.5 F 8.6
14:00 207.2 F 12.7
16:00 217.9 F 5.7
18:00 228.7 F 9.7
20:00 235.5 F 10.9
22:00 239.4 F 11.1
24:00 236.6 F 9.1
26:00 249.6 F 15.3
28:00 260.6 F 21.5
30:00 270.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.2 F 29.1
2:00 272.4 F 29.6
3:00 271.6 F 28.7
4:00 272.3 F 29.2

Min 271.5 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 206.1 F 0.0
10:00 207.5 F 0.0
15:00 202.1 F 0.0
20:00 193.6 F 0.0
25:00 192.9 F 0.0
30:00 189.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:46
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:08:56
CYCLE COMPLETE

04/27/2023 08:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____