
Midmark M11 - v1.0.5
Total Cycles: 3940

Name: RITTER2
Sterilizer ID: V2313228
09/12/2024 05:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.5 F 0.0
2:00 177.3 F 3.6
4:00 201.6 F 8.7
6:00 220.8 F 9.7
8:00 229.8 F 9.9
10:00 233.5 F 8.5
12:00 241.5 F 11.4
14:00 246.3 F 13.9
16:00 259.3 F 20.9
18:00 270.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.7 F 29.7
2:00 271.6 F 29.0
3:00 271.2 F 29.1
4:00 272.4 F 29.7

Min 271.0 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 219.7 F 0.0
10:00 223.6 F 0.0
15:00 213.4 F 0.0
20:00 199.1 F 0.0
25:00 193.2 F 0.0
30:00 190.2 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

09/12/2024 06:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____