
Midmark M11 - v1.0.5
Total Cycles: 2683

Name: RITTER2
Sterilizer ID: V2313228
07/21/2023 06:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.9 F 0.0
2:00 137.8 F 0.0
4:00 155.6 F 1.3
6:00 182.1 F 5.0
8:00 200.0 F 8.9
10:00 214.9 F 13.1
12:00 223.2 F 7.0
14:00 234.0 F 11.5
16:00 233.5 F 8.7
18:00 241.0 F 11.3
20:00 245.6 F 13.5
22:00 257.3 F 19.8
24:00 268.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.7 F 29.5
2:00 271.8 F 29.2
3:00 272.7 F 29.6
4:00 271.7 F 28.9

Min 271.7 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 200.0 F 0.0
10:00 215.5 F 0.0
15:00 214.5 F 0.0
20:00 211.0 F 0.0
25:00 207.6 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:03:44
CYCLE COMPLETE

07/21/2023 07:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____