
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006395
8:09 02/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:38
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 224kPa 4:08
136.3 C 219kPa 4:38
136.2 C 221kPa 5:08
136.1 C 221kPa 5:38
136.4 C 219kPa 6:08
136.4 C 221kPa 6:38
136.4 C 217kPa 7:08
136.4 C 217kPa 7:08
VENTING 7:09
AIR DRYING 7:57
CYCLE COMPLETE 30:57

Digital Signature #
DFEA063064410A80
