
Midmark M11 - v1.0.0
Total Cycles: 4963

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2024 05:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 167.1 F 1.4
4:00 195.1 F 5.9
6:00 214.8 F 11.8
8:00 225.8 F 6.2
10:00 236.3 F 11.4
12:00 234.4 F 8.3
14:00 241.3 F 10.7
16:00 246.4 F 13.2
18:00 258.9 F 20.0
20:00 270.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.4 F 28.5
2:00 271.7 F 28.4
3:00 272.2 F 28.4
4:00 271.7 F 28.5

Min 271.2 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 244.4 F 0.0
10:00 226.2 F 0.0
15:00 230.8 F 0.0
20:00 220.7 F 0.0
25:00 236.6 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

10/03/2024 06:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____