
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008453
12:57 25/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:21
136.0 C 220kPa 4:21
136.4 C 223kPa 4:51
136.4 C 225kPa 5:21
136.2 C 218kPa 5:51
136.4 C 223kPa 6:21
136.4 C 223kPa 6:51
136.4 C 224kPa 7:21
136.3 C 220kPa 7:51
136.3 C 220kPa 7:51
VENTING 7:51
AIR DRYING 8:42
CYCLE COMPLETE 33:40

Digital Signature #
65F6EAC2425DF29A
