
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003202
13:00 23/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:03
STERILIZING 4:05
136.4 C 221kPa 4:05
136.4 C 225kPa 4:35
136.3 C 225kPa 5:05
136.4 C 225kPa 5:35
136.3 C 224kPa 6:05
136.4 C 222kPa 6:35
136.4 C 225kPa 7:05
136.1 C 222kPa 7:35
136.1 C 222kPa 7:35
VENTING 7:35
AIR DRYING 8:27
CYCLE COMPLETE 8:46

Digital Signature #
E0BFB4085A67A3B7
