
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000140
12:26 11/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:59
STERILIZING 4:22
136.0 C 222kPa 4:22
136.1 C 222kPa 4:52
136.3 C 223kPa 5:22
136.3 C 224kPa 5:52
136.2 C 223kPa 6:22
136.2 C 221kPa 6:52
136.4 C 225kPa 7:22
136.4 C 225kPa 7:52
136.4 C 225kPa 7:52
VENTING 7:52
AIR DRYING 8:45
CYCLE COMPLETE 14:35

Digital Signature #
3E3208F8694C576E
