
Midmark M11 - v1.0.5
Total Cycles: 2116

Name: RITTER2
Sterilizer ID: V2313228
03/03/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 174.6 F 2.4
4:00 177.3 F 6.9
6:00 202.9 F 12.7
8:00 225.2 F 7.9
10:00 236.7 F 9.8
12:00 244.3 F 12.4
14:00 246.6 F 13.6
16:00 260.6 F 21.3
18:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.0 F 28.6
2:00 272.7 F 29.1
3:00 272.3 F 29.2
4:00 272.0 F 28.5

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 220.4 F 0.0
5:00 232.4 F 0.0
Dry Cycle Aborted

FILLING: 1:13
HEATING: 18:20
STERILIZING: 4:00
VENTING: 2:47
DRYING: 5:28
TOTAL CYCLE: 00:31:48

03/03/2023 11:09 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____