
Midmark M11 - v1.0.0
Total Cycles: 1170

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2021 12:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.4 F 0.0
2:00 106.1 F 0.0
4:00 131.2 F 0.1
6:00 162.3 F 1.3
8:00 184.5 F 3.9
10:00 204.9 F 7.3
12:00 217.1 F 11.7
14:00 222.1 F 5.8
16:00 233.8 F 10.6
18:00 233.3 F 7.9
20:00 244.3 F 13.0
22:00 242.7 F 11.5
24:00 254.9 F 17.9
26:00 265.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.4 F 29.1
4:00 272.2 F 28.7

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 218.1 F 0.0
10:00 215.6 F 0.0
15:00 222.0 F 0.0
20:00 234.0 F 0.0
25:00 220.2 F 0.0
30:00 206.7 F 0.0

Drying Complete

FILLING: 1:00
HEATING: 27:44
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:06:08
CYCLE COMPLETE

09/14/2021 02:02 PM
ACCEPTED: X
REJECTED:
Operator: U01