
Midmark M11 - v1.0.5
Total Cycles: 2829

Name: RITTER2
Sterilizer ID: V2313228
09/05/2023 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 190.3 F 2.8
4:00 192.9 F 7.4
6:00 209.2 F 12.6
8:00 226.4 F 8.0
10:00 240.0 F 13.4
12:00 240.1 F 11.6
14:00 239.7 F 10.7
16:00 253.1 F 17.3
18:00 264.1 F 23.9
20:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.5 F 29.6
2:00 272.3 F 30.1
3:00 272.1 F 29.4
4:00 273.1 F 30.0

Min 271.6 F 29.2
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 222.6 F 0.0
10:00 231.7 F 0.0
15:00 230.3 F 0.0
20:00 228.2 F 0.0
25:00 227.6 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

09/05/2023 11:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____