
Midmark M11 - v1.0.0
Total Cycles: 5085

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2024 01:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.7 F 0.0
2:00 158.0 F 0.4
4:00 181.8 F 1.1
6:00 197.2 F 4.6
8:00 216.3 F 8.8
10:00 228.7 F 13.2
12:00 231.5 F 8.9
14:00 240.0 F 10.3
16:00 243.8 F 12.6
18:00 243.2 F 11.6
20:00 256.2 F 18.4
22:00 267.4 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.0
3:00 272.6 F 28.8
4:00 271.9 F 28.2

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 213.6 F 0.0
10:00 207.6 F 0.0
15:00 202.5 F 0.0
20:00 209.3 F 0.0
25:00 215.8 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:05
CYCLE COMPLETE

11/13/2024 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____