
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000902
15:03 03/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 4:17
STERILIZING 5:56
136.4 C 225kPa 5:56
136.4 C 232kPa 6:26
136.4 C 231kPa 6:56
136.3 C 232kPa 7:26
136.3 C 232kPa 7:56
136.2 C 231kPa 8:26
136.4 C 232kPa 8:56
136.4 C 231kPa 9:26
136.4 C 231kPa 9:26
VENTING 9:26
AIR DRYING 10:34
CYCLE COMPLETE 10:37

Digital Signature #
CAB68E89A6F5AE06
