
Midmark M11 - v1.0.5
Total Cycles: 1763

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.3 F 0.0
2:00 189.5 F 2.0
4:00 194.2 F 7.2
6:00 215.0 F 13.0
8:00 229.0 F 9.2
10:00 237.2 F 9.4
12:00 244.4 F 12.5
14:00 245.5 F 13.2
16:00 258.7 F 20.4
18:00 270.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.7 F 29.3
2:00 271.7 F 29.2
3:00 272.2 F 29.4
4:00 272.8 F 29.6

Min 271.5 F 29.2
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 229.9 F 0.0
10:00 225.7 F 0.0
15:00 230.7 F 0.0
20:00 226.8 F 0.0
25:00 228.2 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

11/22/2022 04:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____